

All-Ceramic

Chairside Preparation Guide for IPS e.max®


For IPS e.max® indications

| | | IPS e.max Ceram/ZirPress | IPS e.max ZirCAD | IPS e.max CAD/Press |
|------------------|---------------|-----------------------------|---------------------|------------------------|
| Anterior | Crown | | ✓ | ✓ |
| | Veneer | ✓ | | ✓ |
| | Bridge | | ✓ | ✓ |
| Posterior | Crown | | ✓ | ✓ |
| | Bridge | | ✓ | |

Anterior Chairside Preparation Guide


Full-Coverage Restorations

NEW Anterior Crown Preparation



3-Unit Bridge Restorations



3-Unit Bridge Preparation



Veneers


A medium grit, round-ended, diamond bur is used to remove a uniform thickness of facial enamel by joining the depth-cut grooves.

The diamond bur is angled to bevel back the incisal edge.



Thin Veneers

IPS e.max can be pressed to as thin as 0.3 mm for veneers. If sufficient space is present, IPS e.max can be placed over the existing teeth without the removal of any tooth structure. Depending on the case requirements, however, some teeth may need to be prepared to accommodate for the thickness of the ceramic and to ensure proper contour and emergence profile.



*Please note: For additional masking capabilities and/or layering techniques, further reduction may be necessary. Contact your laboratory for further information.
¹ 1.0 mm only suitable with adhesive cementation.


All-Ceramic

Chairside Preparation Guide for IPS e.max®


Posterior Chairside Preparation Guide

Inlays/Onlays



Inlay Preparation



Onlay Preparation




3-Unit Bridge Restorations 3-Unit Bridge Preparation




Full-Coverage Posterior Crown Preparation

NEW

Adhesive Cementation Preparation



Conventional Cementation Preparation



When layered or pressed ceramic margins are preferred in conjunction with a zirconia framework, enhanced gingival esthetics can be achieved with a definitive 90 degree shoulder preparation.